



[FDAHomePage](#) | [CDRH HomePage](#) | [Search](#) | [A-Z Index](#)

[Questions?](#)



# Mammography

Informationformam mographyfa cilityperson nel,inspect ors,a nd consumers  
aboutt hei mplementationoft heMamm ography Quality StandardsAct of1992(MQS A)

[Home](#) • [About](#) • [Regulations](#) • [Guidance](#) • [Certification/Inspection](#) • [Scorecard](#) • [Reports](#) • [Consumers](#) • [Archives](#) • [Contact Us](#)

## InspectionNewsUpdates

- [CommunicationofMa mmographyResultstoReferringHealthcareProvidersa nd Patients](#)(February4,2008)
- [QualityControlTesti ngfor PrintersandMonitors](#) (February 4, 2008)
- [MedicalPhysicistRe sponsibilitiesR egardingLase rPrinterMammogra phy EquipmentEvaluation\(M EE\)](#)( November 27,2007)
- [AttentionMam mographyC ommunity -PendingInspectionSoftw areCita tions willbeActivatedonJa nuary1,2008](#) (October31,2007)

## CommunicationofMammography Res ultstoReferri ngHealthcare Providersan dP atients

TheDivisionofMamm ographyQ ualityandR adiationProgramsha sbe enmadeaw are that sometimesthereareissueswiththe m methodssomef acilitiesus etotra ckthat mammographyresultsare provide dtore ferringhealthcareproviders andpatie nts.Some MQSAfac ilitiesmaynot be consiste ntlytrackingormonitoringtha tmedicalreports and/orlaysumm ariesar ebeingpr ovidedontimeor atall.Asar esult,somereferr ing healthcareprovide rsandpa tientsare notre ceivingtheirexam resultsorare notrece iving theminatimelymanner.I nm ostinstances,facilitieshave establishedwrittenpr ocedures inplacet oissuethere ports,howe ver, inpractice,thetrac kingofre portsissometimes not adequateto ensuretimize lyissuanceof themammogr aphyresults.

MQSAallowsfacilitiestodeve loporuseaprocedurea ndtracking sys temthatwor ks best fort hem.Facilitie sne edtomonitortheirsy stemstoe nsurethatthe irpolicyand proceduresareactuallya ndc orrectlybe ingfollowe d.FD As upportstheuseofcomputer

trackingandpaperorpatientlogssystemstoassistintrackingtimelinessofmedical reportsandlaysumaries.Someradiologycomputerreportingsystemsca ntrack individualreport sandgenesumaryreportsindicatingwhenamammographyreport orlaysummaryhasbeenissued.By routinelycheckingthesesummaryreports, facilities can ensurethat allmammographyresults havebeensubmittedin a timely manner.Facilities areencouragedtocheckwiththeir computersupportvendorstoensureiftheir softwarecan generatesuchsumaryreports.FDA encourageallfacilities to ensure the current systems, whether they are computerized or paper log stored, ensure that they communicate mammography results to all patients and the referring health care providers.

Another sometimes overlooked problem deals with facilities that fax or mail medical reports and/or laysumaries. Policies and procedures to deal with faxes and emails that are reported as “failed” are important to ensure that reports and laysumaries are resent in a timely manner.

Facilities may obtain additional information about medical records and mammography reports from the MQSA Policy Guidance Help System located on this website (Search on keywords: Medical Records, Mammography Reports, Mammographic Assessments, Written Report, Record keeping, and Communication of Mammography Results.)

## **Quality Control Testing for Printers and Monitors**

### **1. Printer and/or Monitors without QC Manuals**

Under the Mammography Quality Standards Act (MQSA) final regulation 900.12(e)(6), facilities using a mammographic modality other than screen-film must follow a quality control program that is “substantially the same as the quality assurance program recommended by the image receptor manufacturer...”. While all full-field digital mammography (FFDM) manufacturers have quality control (QC) manuals, in some cases, the QC manual instructs the facility to test monitors and printers according to the component’s QC manual. In these cases, it is the responsibility of each facility to ensure that it obtains and follows the component’s QC manual for its monitors and printers.

### **2. Facilities Using the Same Printer/Monitor with FFDM Units from Different Manufacturers**

For facilities that are using FFDM units from different manufacturers, each with its own QC requirements for printers and monitors, there may be uncertainty regarding the QC tests to perform on the components. The following three examples should help facilities decide.

- a. Each FFDM manufacturer QC manual requires that the same or equivalent test be done, but at different time frequencies. In this case, facilities need to perform the test at the more stringent frequency.
- b. Each FFDM manufacturer QC manual requires that different but equivalent tests be done. In this case, facilities may perform only one of the tests at the more stringent frequency. The medical physicist should provide a written statement for the facility’s

quality control records, indicating that in his or her opinion, the two tests are equivalent. Each FFDM manufacturer QC manual requires that different tests (note equivalent) be done. In this case facilities need to perform each test at the frequency required in the respective FFDM manufacturer QC manual.

## Medical Physicist Responsibilities Regarding Laser Printer Mammography Equipment Evaluation (MEE)

We recently learned that some medical physicists are not including an appropriate testing of the laser printer in the survey/MEE report to the facility. There seems to be confusion between the required MEE testing of a laser printer: (1) before it is first put into clinical use; (2) following reassembly; or (3) after major repairs vs. the routine Quality Control (QC) testing of the laser printer after it is placed into clinical use. In some cases, the medical physicists did not test the laser printer at all. In other cases, they did inappropriate testing and, in a few cases, they did appropriate testing but did not include documentation of the tests in the facility survey/MEE report. Therefore, we want to clarify the MQS requirements regarding this issue.

The final regulations require that all test procedures be conducted as specified in the QC manual of the FFDM system manufacturer. Some of these QC manuals specify both periodic QC (daily, weekly, or annual) and MEE testing for the laser printer. Other manuals do not address the subject and refer the user to the laser printer manufacturer's QC manual. For easy reference, the following table summarizes laser printer test procedures in current QC manuals:

FFDM System	QC Manual	Weekly/Daily	Annual	MEE	QC Procedures - Comments
GE-All systems	All	Yes*	No	No	*Per printermfr.QC manual
Fischer	Rev.10-10/07	Daily check	No	No	Follow printermfr.QC manual
Selenia	Rev.7 –8/07, **Sec.2.1	Yes	No	Yes**	Follow the Selenia QC manual
Siemens	Rev.5 –4/07	Before clinical use	No	Yes	Follow printermfr.QC manual
Fuji	3rd Edit. –4/07	Yes	Yes	Yes	Follow applicable printer QC manual

As the above table shows, some of the FFDM QC manuals do not specifically address the laser printer MEE testing requirements when first installed, reassembled, or after having undergone a major repair. Hence, the facility or medical physicist has to obtain this information from the laser printer manufacturer. In some cases, the QC manuals only address the interface between the FFDM unit and the laser printer. They do not address the basic requirement that the laser printer, before it is used clinically for mammography, has to be operating and signed by the laser printer manufacturer. Since the MEE requires

that both these are as been checked, the medical physicist may have to consult both the FFDM manufacturer's QC manual and the laser printer QC operator's manual to determine which tests are required to assure that the laser printer is functioning properly.

Some medical physicists incorrectly assume that simply scoring a laser printed phantom image satisfies all the requirements of the laser printer MEE. However, in most cases, this practice would not be acceptable. To avoid unnecessary follow-ups by MQSA inspectors, we urge all medical physicists to review the FFDM manufacturer's QC manual and, where necessary, the laser printer QC operator's manual to determine the appropriate testing. The yearly need to clearly document the testing of the laser printer in their reports.

### **Attention Mammography Community - Pending Inspection Software Citations will be Active on January 1, 2008**

On January 1, 2008, FDA will modify the current inspection software (FIS S Version 6.03) to allow citations for failures in the following areas:

- Failing phantom score(s) in full field digital mammography (FFDM) systems – As you know, MQSA inspectors have been scoring phantom image sin FFDM systems since 2006, but we have withheld issuing citations for failing these image scores to give the inspectors a decent time to become comfortable with the FFDM scoring procedure and to get the facilities a “heads-up” view of four intent in this area. Effective January 1, 2008, failing phantom image scores in FFDM systems will have the same observation level as screen-film (S-F) systems.
- Failing to take corrective action before resuming clinical use if the dose reported by the medical physicist exceeds 3 mGy (300 mrad). Specifically, we added the following data entries in early 2006:
  - Dose value (mrad) reported -----, and if the inspector entered a value over 300 mrad, the software prompts the inspector to answer the following question:
  - C/A taken before resuming clinical use? (yes/no)

Thus, effective January 1, 2008, if the inspector answers “no” to the above question, the software program will issue a level 2 observation, which will be applicable to both S-F and FFDM systems.

Updated February 5, 2008



[FDAHomePage](#) | [CDRH HomePage](#) | [Search](#) | [A-Z Index](#)


[Questions?](#)



# Mammography

Informationformam mographyfa cilityperson nel,inspect ors,a nd consumers  
aboutthei mplementationofthe MammographyQualityStandard sActof19 92(MQS A)

[Home](#) • [About](#) • [Regulations](#) • [Guidance](#) • [Certification/Inspection](#) • [Scorecard](#) • [Reports](#) • [Consumers](#) • [Archives](#) • [Contact Us](#)




## MQSA Facility Score Card



☒ **UpdatedTrends in  
MammographyDoseandImage  
Quality**

(Related Article: [DoseandImage  
QualityinM ammography:  
Trends duringtheFi rstDecade  
ofMQSA](#) )

**DavidC.Speli c,Ph .D.**  
USF oodan dDr ug Administration  
Divisiono fM ammographyQualityandRadiation  
Programs  
1350Pic cardD riveHFZ -240  
Rockville,MD20 850  
Phone:301 -594-3525  
Email: [dcscdrh.fda.gov](mailto:dcscdrh.fda.gov)  
6/16/2006



This graphic displays a average value of mean glandular dose and estimates of image quality in mammography for the period from the early 1970's to 2005. Dose in mammography have consistently decreased with time, with the most substantial reductions in dose occurring from the early 1970's to the early 1980's. Image quality data is presented from the mid-1980's to present, and shows consistent improvement with time.

The early studies of 1974 and 1976 reported substantial contributions from xeromammography and direct-film procedures. Even during the first year of MQSA inspections in 1995, there were a number of facilities still using xeromammography. Technological improvements such as efficient screen-film technologies, molybdenum target-based x-ray units for mammography, and improvement in film processing quality contributed to further reduction in dose during the eighties and into the nineties. Dose and image quality trends have leveled off in recent years as the technical aspects of mammography became optimized. For further discussion on trends in dose and image quality in mammography, refer to [the article posted on the FDA's website](#).

Average values for mean glandular dose are either reported by, or derived from data reported by the following sources:

1974	Bicehouse HJ. Survey of Mammographic Exposure Levels and Techniques Used in Eastern Pennsylvania. 7th Annual National Conference on Radiation Control, 1975. DHEW Publication (FDA) 76-8026.
1976	Butler PF, Jensen JE. Breast Exposure: Nationwide Trends; AMammographic Quality Assurance Program-Results to Date. Radiologic Technology 50(3),

	1978;pp251 -257.
1980	BreastExposure:Na tionwideTr ends.In:Internalproje ctpro gress report. RockvilleMD:Bureauof RadiologicalHea lth,USD epartmentofHealth and HumanServices,FoodandDru gAdministration, 1981.
1985, 1988, 1992	ConwayB J,S uleimanOH, Ruete rFG, AntonsenRG ,Slayton RJ.National SurveyofMam mographic Facilitiesin1985,1988,a nd1992.Radiology 1994;191:323 -330.
1995to present	MammographyQuality StandardsAc t (MQSA)inspec tionfindings.

ImageQualityscoresare reportedfr omthe follow ingsour ces:

NOTE:All imagequa lityscoresare repo rtedwithoutartifactsubtr action.

1985	RMI152phantom with'C 'insert;1985Nationwide EvaluationofX -ray Trends(NEXT)surveyda ta.
1988	RMI156pha ntomwith'C 'insert;1988 NEXTsurv eydata.
1992	RMI156pha ntomw ith'D'insert ;1992NEXTsurvey data.
1995to present	MammographyQualityStanda rdsA ct(MQSA) inspectionfindings

UpdatedJune16,200 6



[FDAHomePage](#) | [CDRH HomePage](#) | [Search](#) | [A-Z Index](#)

[Questions?](#)



# Mammography

Informationformam mographyfa cilityperson nel,inspect ors,a nd consumers aboutthei mplementationoft heMa mmographyQualityStandard sActof19 92(MQSA)

[Home](#) • [About](#) • [Regulations](#) • [Guidance](#) • [Certification/Inspection](#) • [Scorecard](#) • [Reports](#) • [Consumers](#) • [Archives](#) • [Contact Us](#)



## MQSA Facility Score Card



### MQS ANat ionalSta tistics

Inthisse ctionofth eMQSASc ore card,we presentt he most commonlyre quested nationalstatisticsr egardingtheM QSA program.Thesestatistic sa re updatedonth e firstofeachmonth.

Certifiedfacili ties,asofOc tober1,2007	8,837
---	-------

Certificationst atistics,a sof Ma rch 1,20 08

Totalcertifiedfacilities/Totala ccreditedunits	8,871/13,547
--	--------------

Certifiedfacilitie swithFFDM <sup>2</sup> units/AccreditedFFD Munits	2,847/4,31 0
--	--------------

FY2008inspectionsta tistics,a sofM arch1,2008

Facilitiesinspe cted	3,284
----------------------	-------

Total unitsatinspecte dfa cilities	4,917
------------------------------------	-------

Percentofinspec tionswhe rethe highestnonc ompliance wasa:

Level1violation	1.3%
-----------------	------

Level2viola tion	16.1%
------------------	-------

Level3violation	6.1%
-----------------	------

Percentofinspec tionswithnoviolation	76.5%
--------------------------------------	-------

Totalannu almmammographypr oceduresreporte d,a sofMarch1,2008 <sup>1</sup>	35,787,362
--	------------

<sup>1</sup>Thisnumberis ana ggregateofth etalnumberofproceduresperformedannu ally asr eportedbyfacili tieso theiraccreditationbodies. Facilitiesareaskedtodisclosethisi nformationat theirinitialaccredit ation,andt hena tthetime of theirre -accreditation,whichtakes p lace onceeverythreeyea rs. FDA begancollectingthesedatain 1998.Theaggregatedoesnotreflectthe currennumber ofprocedure sperforme dat thesefacili ties,butonlythenum bers reported bythemduringthe three-year periodprior othe cur rentdate.We have aggregatedonlythe numbers reportedbycertif ied,n on-VeteransAdminis trationfacili ties.

<sup>2</sup>FFDM -Full FieldDigitalMammographyunit.

UpdatedMa rch 3,2008