

Proton Therapy QA & Operations Loma Linda University Medical Center

Baldev Patyal, Ph.D., Chief Physicist
Department of Radiation Medicine



LOMA LINDA UNIVERSITY
MEDICAL CENTER

AAPM-2013_Patyal

Facility Overview

&

Operations



LOMA LINDA UNIVERSITY
MEDICAL CENTER

AAPM-2013_Patyal


James M. Slater, MD, Proton Therapy and Research Center

World's first hospital based proton treatment facility (1990)

- Accelerator: a **synchrotron**
- Passive beam scattering
- **Four** treatment rooms, **Five** treatment beam lines
 - three isocentric gantries
 - one with robotic patient positioner
 - one horizontal beam line
 - one fixed eye beam line
- One research room (3 horizontal beam lines)
- Variable energy capability with energies up to 250 MeV
- Accelerator runs 6 days a week, 24 hrs. a day
- One day per week for repair and preventive maintenance
- Active beam scanning (**ABS**) under development



Proton Treatment Facility Layout



Typical Proton Facility Weekly Usage

<u>MODE</u>	<u>HOURS</u>	<u>PERCENTAGE</u>
Treatment	94	56%
Calibration	24	14%
Maintenance	12	7%
Research	10	6%
Upgrades	16	10%
Operations	5	3%
Other	7	4%
TOTAL	168	100%



Timeline of a Proton Treatment @LLUMC

<u>Procedure</u>	<u>Time Needed</u>	<u>Time before Tx</u>
Patient orientation & education	1-2 hours	7-10 days
Patient immobilization & imaging	15-30 min	7-10 days
Treatment planning & plan QA	1- 3 days	5-8 days
Device manufacture & calibration	1 day	1-3 days
Treatment session	15-45 min
Patient setup	5-10 min
Alignment verification & approval	5-25 min
Treatment	2-5 min
Room reset	2-5 min



Time Requirements for Routine QA Tasks

<u>Task</u>	<u>Time Needed</u>	<u>Comments</u>
Simple calibration	10 min/field	5 min/additional field
Complex calibration	15-20 min/field	10 min/additional field
Problem fields	30-60 min/field	Small, odd shaped fields
Radio surgery fields	20-30 min/field	10-15 min/additional field
Eye-beam fields	45 min to 2 hours	Small mod-wheels, range
Model calibration	5-7min/field	Does not need beam
Device checks	5 min/device	Boluses and apertures
Daily QA	15 min/energy 20 min imaging sys lasers, table etc.	
Calibration checks	15-20 min/field	




Staffing

- Staff cross-trained in both x-ray and proton therapy
- 5 Ph.D. physicists, 5 M.S. physicists, 10 dosimetrists (9 physicians)
- 2 proton calibration physicists (night shift)
- 4 proton treatment rooms, 4 linear accelerators, one CT-Sim, 2 CT scanners and brachy service
- EMR: ARIA/Oncochart
- Three academic programs
- Ongoing research & development (**without treatment interruption**)
- Physics workload: protons to x-rays: 65:35



Patient Mix by Diagnosis (since inception)



17,288 Patients treated through 06/2013




110 patients a day (year-to-date)

Quality Assurance



LOMA LINDA UNIVERSITY
MEDICAL CENTER

A Proton Treatment System



1 Accelerator

- Beam position monitoring
- Beam energy monitoring
- Beam current monitoring
- Beam steering

2 Beam Transport System

- Beam routing to tx rooms
- Beam profile and centering

3 Nozzle

- Beam shaping/steering
- Beam monitoring
- Dose monitoring

4 Gantry

- Beam rotation/aiming
- Imaging

5 Patient Positioner



LOMA LINDA UNIVERSITY
MEDICAL CENTER

Daily QA Checks

- Review beam delivery system logbook
- Verify integrity of the control system database for each beam-line
- Inspect treatment room (motion locks, snout extension, snout motion, range of table motions, hand pendent function etc.)
- Inspect treatment console area (indicator lamps, token key, backup monitor reset, audio-visual patient monitors)
- Check localization lasers (2 mm)
- Verify modulator propeller interlock
- Verify functionality of barcode scanning system
- Check the functionality of the area radiation monitor within the room
- Run the daily DI tests
- Perform a daily QA for each energy
- Perform patient calibration for each new portal



Daily Proton Beam QA

- A procedure to standardize operating conditions for each day, each room, each energy
- Daily beam calibration performed under **standard conditions** for each energy
 - Isocenter at Center of modulation (6 cm mod wheel)
 - Nozzle extension
 - Standard aperture
- Detector response compared with standard calibration (Pass/Fail)
 - Calibration factor used for patient portal calibration
 - Also includes verification of **entrance and distal dose and range verification**


Daily beam calibration summary report

Daily Calibration summary		Run Number: 55653				
Portal ID:	AB					
Outcome:	Finished					
Standard Calibration #:	103					
Temperature (deg C.):	19.40					
Pressure (mm Hg.):	735.40					
Beam Energy (Mev.):	155					
Counts:	1000000					
Modulator Wheel:	060					
Range Shifer ID:	0					
Nozzle Extension:	48.29					
Target Identifier:	wpc 88					
Validation:	VALID					
Comment:	Combo detector passed daily calibration.TIC 3 detector passed daily calibration.					
Sequence Number	Channel	Background Factor	Raw Counts	Corrected Counts	Monitor Units	Calibration Factor
Detector: Combo						
1	SEM	3.64	7294	6826.909	6826.91	48.4341
1	TIC 1 SF	3.14	666510	666107.062	666107.06	4725.7578
1	TIC 1 Q1	3.11	164528	164128.922	164128.92	1164.4277
1	TIC 1 Q2	1.74	155416	155192.719	155192.72	1101.0289
1	TIC 1 Q3	3.11	164893	164493.922	164493.92	1167.0172
1	TIC 1 Q4	2.69	154578	154232.812	154232.81	1094.2189
1	TIC 2 Q1	3.00	157415	157030.031	157030.03	1114.0640
1	TIC 2 Q2	3.16	158365	157959.500	157959.50	1120.6582
1	TIC 2 Q3	3.11	153664	153264.922	153264.92	1087.3521
1	TIC 2 Q4	3.11	152639	152239.922	152239.92	1080.0801
Detector: TIC 3						
1	TIC 3 CPP	3.04	3908	3517.896	3517.90	24.9580
1	Backup	0.32	19676	19634.936	19634.94	139.3018
1	TIC 3 R1	3.15	1004275	1003870.750	1003870.75	7122.0532
1	TIC 3 R2	3.25	623925	623507.938	623507.94	4423.5342
Detector: Ion Chamber						
1	0	0.00	0	1.233	140.95	114.3167
Ratios:			Numerator	Denominator	Value	
TIC 1 Up/Down			59487.50	638048.38	0.09323	
TIC 1 Left/Right			132487.50	638048.38	0.20764	
TIC 2 Up/Down			948468.75	620494.38	1.52857	
TIC 2 Left/Right			195450.00	620494.38	0.31499	
TIC 3 Up/Down			-23938.28	86053.50	-0.27818	
TIC 3 Left/Right			-20031.64	86731.70	-0.23096	
TIC 3 R1/SEM			1003870.75	6826.91	147.04617	
TIC 3 R1/TIC 3 CPP			1003870.75	3517.90	285.36108	
TIC 3 R1/TIC 1 SF			1003870.75	666107.06	1.50707	
TIC 1 SF/SEM			666107.06	6826.91	97.57082	



Patient Specific QA: Each New Portal

- Check aperture for each new portal
- Check compensator bolus
 - Visual inspection
 - Probe plunge depths at select points against treatment planning data
- Correct barcodes for devices
- Perform patient specific calibration (model or physical)
- Do a second person check of patient calibration before the first treatment



Patient Specific QA: Portal Calibration

Physical Calibration

- Performed with portal specific energy and patient devices
- Ion chamber placed at an equivalent depth in a solid phantom
- Prescribed dose delivered and output measured
- The counts, output and calibration factors generated for the detectors are stored in a patient file
- The beam calibration performed the day of treatment is used for portal calibration
- Model calibrations used when possible
- Small fields, large air gap, thick bolus: require physical calibration

Patient physical calibration report

Data for patient set-up file:		Data for physical calibration:	
Physical dose (cGy)	163.6 ✓	Iso to cal. pt. distance (cm)	0.0
Energy (MeV)	249.5 ✓	Calibration depth (cm)	20.61 ✓
Modulator wheel ID	80 ✓	Blocks used:	3SPMSA 20.657
Range shift	0 ✓		
Nozzle extension	29.1 ✓		
Table angle (deg)	-90 ✓		
Effective f. s. at iso (cm)	5.52 ✓		
Cone size (cm)	15 ✓		

3-point bolus check:					Overlay checks:	
Point location	Coordinates (x : y)	WEQ. thickness (cm)	Phys. thickness (cm)	Meas. phys. thickness (cm)	Bolus	
Minimum	(-0.87;-1.53) ✓	6.67 ✓	6.80 ✓	6.80 ✓	AA 1/16/11	
Calibration	(0.25-0.39) ✓	7.08 ✓	7.22 ✓	7.22 ✓	Aperture	
Maximum	(-1.87;-2.57) ✓	11.53 ✓	11.75 ✓	11.76 ✓	AA 1/16/11	

Center of modulation calculation:		Beam range check:	
Beam range (cm)	31.50	Beam range (cm)	31.50
- modulation + 2 (cm)	08.23 / 2 ✓	- max. treatment depth (cm)	24.83 ✓
- bolus thickness cal. pt. (cm)	07.08 ✓	- min. bolus thickness (cm)	06.67 ✓
= COM depth (cm)	20.31 ✓	= excess range (cm)	00.00 ✓

SMGD = 7.114

Manual calculation of backup monitor:			
Backup =	$C_{mod} \times C_{cor} \times C_{oz} \times phys. dose \times K_{app}$		
=	1.101 × 1.054 × 0.998 × 163.6 × 121.6776		
=	23054 ✓		

Physical calibration output:		General calibration output:	
Ion chamber ID	wpc 88	Run ID	51706
IC calibration factor	111.60	Date	01/06/2011
Temperature (°C)	19.7	Time	2:27
Pressure (mm Hg)	734.6	Saved? <input checked="" type="checkbox"/>	Gantry set to
Nozzle extension used	29.0 ✓		
Requested dose (cGy)	163.6		
Delivered dose (cGy)	163.6756		
IC charge (nC)	1.4289		
IC dose (nC × calfactor × K _{TP})	163.6927		
Time used (min.)	2.2		
Primary TIC MU	—		
Norm. primary TIC MU	—		
Backup TIC MU	22848		
Norm. backup TIC MU	22853 ✓		
Difference w.r.t. manual calc.	0.952 ✓		

Treatment settings:	
Backup setting:	Chamber at (25, 0.4)
= backup TIC MU × 1.1	Bolus replaced by SMGD.
=	25121 ✓
Timer setting (min):	
- run time × 1.5	
=	3.3 ✓

Planner	JA 1-3-11	Plan QA	SL 1/6/11
Calibrator	W. O. K. 1/7/11	Cal Check	W. O. K. 1/7/11

Rev. 2.6, 06/20/2000

This sheet printed 01/03/11, 6:29 PM

JAN 06 2011



Monthly QA

- Review daily QA records
- Verify the room secure and other interlocks
- Verify the backup dose monitor functions properly
- Check snout extension accuracy at three extensions
- Beam pause and reset functions
- Verify the integrity and alignment of the scatterers
- Verify the patient calibration system
- Emergency off switches
- Beam flatness and symmetry
- X-ray and proton beam centering
- DI calibration tests



Annual QA...

Safety Checks

- Facility shutdown switches
- Area shutdown switches
- All beam safety interlocks
 - Beam energy interlock
 - Beam centering interlock
 - Backup counter interlock
 - Modulator wheel interlock
 - Door interlock
- Room secure
- Beam on indicator
- Beam pause and resume functions
- Calibrate area radiation monitors
- Audio-visual system function
- Radiation warning signs

Mechanical Checks

- Gantry angle readout accuracy
- Patient positioner readout accuracy
- Patient positioner sag
- Snout extension readout accuracy
- Laser alignment accuracy
- Modulator wheel visual inspection
- Inspection of block and bolus doors and latches in the nozzle
- Hand pendant operation (also daily)

X-rays & Imaging Checks

- Image magnification accuracy
- kVp and mAs accuracy
- Image quality



Annual QA...

Proton Checks

- Proton vs. X-ray field centering
- Location of the effective source
- Location of the virtual source
- Linearity of dose per monitor unit for the primary channel

Proton Dosimetry Checks

- Field size dependent factors for all clinically used energies ($\pm 2\%$)
- Modulation factors ($\pm 2\%$)
- Depth dose profile
- Dose per monitor unit for the primary and all backup channels at selected gantry angles ($\pm 1\%$)
- Bolus gap factors
- **Calibration protocol: IAEA-TRS-398 (ICRU 78)**

Proton Beam Quality Checks

- Lateral field symmetry
 - energy and gantry angle
- Lateral field flatness
 - energy and gantry angle
- Lateral penumbra widths
 - energy and gantry angle
- Range uniformity (picket fence)

QA Equipment Tests

- Standard ion chambers and electrometers calibration factors
 - Current ADCL calibrations
- In house ion chambers and electrometers calibration factors
 - Periodic inter-comparisons



Summary

- Quality begins at the top! Institution leadership has to show tangible commitment to quality
- To maintain quality and to improve quality, a Quality Management Program; a written document of quality expectations, of procedures and policies, and roles and responsibilities should be developed
- To implement quality programs, the institution should make available the needed resources: **personnel, tools and time**
- The main purpose of a hospital based proton center is to treat patients. But to treat patients right, a thorough QA program is essential. Don't squeeze out QA time to accommodate more patients
- QA procedures and policies should be strictly followed

