




## Features and Weaknesses of Phantoms for CR/DR System Testing

Donald Peck, PhD  
Henry Ford Hospital


## Physics testing of image detectors

- Parameters to test
  - Spatial resolution
  - Contrast resolution
  - Uniformity/geometric distortion
  - Dose response/signal dynamic range
  - Noise
- Experiments/testing methods
  - Direct measurements
  - Phantoms/image tools
    - Qualitative
    - Quantitative






## Phantom Types

- Attenuation
- High Contrast/Spatial Resolution
- Low Contrast/Contrast Sensitivity
- Anthropomorphic




## Attenuation test tools

- Metals
  - Aluminum, Copper, etc.
- Plastics/Composite Materials
  - Lucite, Tissue Equivalent, etc.
- Water




## Detectors - CR Scatter Sensitivity

- Digital detectors can be significantly more sensitive to scatter radiation than traditional phosphor screens
  - Scatter needs to be considered when testing systems




## Attenuator positioning

- Modifying beam quality
  - Position attenuators far from detector to minimize scatter contribution in measurement
- Simulating patient attenuation
  - Position close to detector in same location as patient



### Attenuator Construction

- Attenuator "purity" may not be acceptable for the measurement
  - Measurement of mammography HVL requires attenuators that are at least 99.9% Aluminum
- Tissue equivalent materials may not be uniform



7


### Attenuation test tools

- Easy to use
- Placement of attenuator needs to be considered based on the test
- Purity or Uniformity of material may not be adequate for some tests

8

### High Contrast/Spatial Resolution Test tools


- Line pair patterns



- Mesh patterns




- Edge phantoms




9

### Line pair patterns




10

### Aliasing and Moiré Effect




11

### Line pair patterns



12


## Moiré Effect



13

## MTF/DQE Measurement


- IEC 62220-01 (2003)
  - Method for determining Detective Quantum Efficiency (DQE) of digital imaging systems
  - Defines specifications for a test device required to make these measurements



14

## MTF/DQE Measurement Issues


- Requires Pre-processed image values that are "linear" with exposure
- Determination of edge response
  - Need to bin pixel data along edge
  - Phantom positioning critical for consistent results
- Smoothing/fitting of edge response curves to allow utilization of Fourier Analysis
  - Variations in method used may produce different results
  - Important to standardize if comparing to other MTF/DQE measurements



15

## MTF/DQE Measurement Issues


- Fluorescent radiation
  - Only issue at high kVp
  - Important if comparing to other MTF/DQE measurements
- Noise Power Spectrum (NPS) determination
  - Need to remove effects of trends associate with heel effect, etc.
  - Variations in method used may produce different results
  - Important to standardize if comparing to other MTF/DQE measurements



16

## MTF Measurements

- Quantitative results
- Good indication of changes
- Subtleties in the measurement can make comparisons between measurements by different tests inaccurate



17


## High Contrast/Spatial Resolution Test tools

- Line pair patterns
  - Subjective
- Edge Phantoms
  - Objective
    - MTF Determination
  - Requires development of software to do the calculations
    - Task Group No. 162 "Research Software for 2D Image"
  - Valid for determining if changes have occurred over time if performed "consistently"
  - Requires standardization of methods used if comparisons between systems or results from different physicists are compared

18

## Low Contrast/Contrast Sensitivity test tools

- Contains objects of varying size and attenuation
- Requires observers to determine which objects are visible
  - Subjective




19

## Contrast threshold detection index (TCDD)

- TCDD gives an indication of the lowest contrast detectable ( $C_T$ ) as a function of the detail size (the square root of the detail Area,  $A$ ) and can be quoted in terms of the threshold detection index ( $H_T$ )

$$H_T(A) = 1/[C_T \cdot A^{1/2}]$$

- High value for  $H_T(A)$  indicates good visibility



20

## Institute of Physics and Engineering in Medicine (IPeM)

- Goals:
  - Improving standards in clinical practice
  - Providing advice on scientific and engineering issues in healthcare to other healthcare professionals, *government* and the public.
- Develops Reports and other publications to achieve these goals
  - Owns several journals:
    - Physics in Medicine and Biology
    - Physiological Measurement
    - Medical Engineering and Physics
  - Report 91 Recommended Standards for the Routine Performance Testing of Diagnostic X-Ray Imaging Systems
    - Specifies the use of phantoms throughout the testing procedures

GROUP TEST OBJECT	IPeM Report No. 91 PAGE No.	DETAILS	REF
TCDD	45	Low contrast resolution	6.1.1.1.1.1
+	45	Low contrast sensitivity	6.1.1.1.1.2
+	45	Limiting spatial resolution	6.1.1.1.1.3
+	45	Limiting spatial resolution	6.1.1.1.1.4
TCDD	46	Scaling errors	6.1.1.1.1.5
+	46	Scaling errors	6.1.1.1.1.6
TCDD and NBSA	47	High contrast resolution	6.1.1.1.1.7
+	47	Threshold contrast	6.1.1.1.1.8
+	47	High contrast resolution	6.1.1.1.1.9
+	47	Threshold contrast	6.1.1.1.1.10
TCDD and NBSA	48	Baseline spatial resolution	6.1.1.1.1.11
+	48	Baseline spatial resolution	6.1.1.1.1.12
TCDD and NBSA	49	Baseline spatial resolution	6.1.1.1.1.13
+	49	Baseline spatial resolution	6.1.1.1.1.14
TCDD and NBSA	50	Baseline spatial resolution	6.1.1.1.1.15
+	50	Baseline spatial resolution	6.1.1.1.1.16
TCDD and NBSA	51	Baseline spatial resolution	6.1.1.1.1.17
+	51	Baseline spatial resolution	6.1.1.1.1.18
TCDD and NBSA	52	Baseline spatial resolution	6.1.1.1.1.19
+	52	Baseline spatial resolution	6.1.1.1.1.20
TCDD and NBSA	53	Baseline spatial resolution	6.1.1.1.1.21
+	53	Baseline spatial resolution	6.1.1.1.1.22
TCDD and NBSA	54	Baseline spatial resolution	6.1.1.1.1.23
+	54	Baseline spatial resolution	6.1.1.1.1.24
TCDD and NBSA	55	Baseline spatial resolution	6.1.1.1.1.25
+	55	Baseline spatial resolution	6.1.1.1.1.26
TCDD and NBSA	56	Baseline spatial resolution	6.1.1.1.1.27
+	56	Baseline spatial resolution	6.1.1.1.1.28
TCDD and NBSA	57	Baseline spatial resolution	6.1.1.1.1.29
+	57	Baseline spatial resolution	6.1.1.1.1.30
TCDD and NBSA	58	Baseline spatial resolution	6.1.1.1.1.31
+	58	Baseline spatial resolution	6.1.1.1.1.32
TCDD and NBSA	59	Baseline spatial resolution	6.1.1.1.1.33
+	59	Baseline spatial resolution	6.1.1.1.1.34
TCDD and NBSA	60	Baseline spatial resolution	6.1.1.1.1.35
+	60	Baseline spatial resolution	6.1.1.1.1.36
TCDD and NBSA	61	Baseline spatial resolution	6.1.1.1.1.37
+	61	Baseline spatial resolution	6.1.1.1.1.38
TCDD and NBSA	62	Baseline spatial resolution	6.1.1.1.1.39
+	62	Baseline spatial resolution	6.1.1.1.1.40
TCDD and NBSA	63	Baseline spatial resolution	6.1.1.1.1.41
+	63	Baseline spatial resolution	6.1.1.1.1.42
TCDD and NBSA	64	Baseline spatial resolution	6.1.1.1.1.43
+	64	Baseline spatial resolution	6.1.1.1.1.44
TCDD and NBSA	65	Baseline spatial resolution	6.1.1.1.1.45
+	65	Baseline spatial resolution	6.1.1.1.1.46
TCDD and NBSA	66	Baseline spatial resolution	6.1.1.1.1.47
+	66	Baseline spatial resolution	6.1.1.1.1.48
TCDD and NBSA	67	Baseline spatial resolution	6.1.1.1.1.49
+	67	Baseline spatial resolution	6.1.1.1.1.50
TCDD and NBSA	68	Baseline spatial resolution	6.1.1.1.1.51
+	68	Baseline spatial resolution	6.1.1.1.1.52
TCDD and NBSA	69	Baseline spatial resolution	6.1.1.1.1.53
+	69	Baseline spatial resolution	6.1.1.1.1.54
TCDD and NBSA	70	Baseline spatial resolution	6.1.1.1.1.55
+	70	Baseline spatial resolution	6.1.1.1.1.56
TCDD and NBSA	71	Baseline spatial resolution	6.1.1.1.1.57
+	71	Baseline spatial resolution	6.1.1.1.1.58
TCDD and NBSA	72	Baseline spatial resolution	6.1.1.1.1.59
+	72	Baseline spatial resolution	6.1.1.1.1.60
TCDD and NBSA	73	Baseline spatial resolution	6.1.1.1.1.61
+	73	Baseline spatial resolution	6.1.1.1.1.62
TCDD and NBSA	74	Baseline spatial resolution	6.1.1.1.1.63
+	74	Baseline spatial resolution	6.1.1.1.1.64
TCDD and NBSA	75	Baseline spatial resolution	6.1.1.1.1.65
+	75	Baseline spatial resolution	6.1.1.1.1.66
TCDD and NBSA	76	Baseline spatial resolution	6.1.1.1.1.67
+	76	Baseline spatial resolution	6.1.1.1.1.68
TCDD and NBSA	77	Baseline spatial resolution	6.1.1.1.1.69
+	77	Baseline spatial resolution	6.1.1.1.1.70
TCDD and NBSA	78	Baseline spatial resolution	6.1.1.1.1.71
+	78	Baseline spatial resolution	6.1.1.1.1.72
TCDD and NBSA	79	Baseline spatial resolution	6.1.1.1.1.73
+	79	Baseline spatial resolution	6.1.1.1.1.74
TCDD and NBSA	80	Baseline spatial resolution	6.1.1.1.1.75
+	80	Baseline spatial resolution	6.1.1.1.1.76
TCDD and NBSA	81	Baseline spatial resolution	6.1.1.1.1.77
+	81	Baseline spatial resolution	6.1.1.1.1.78
TCDD and NBSA	82	Baseline spatial resolution	6.1.1.1.1.79
+	82	Baseline spatial resolution	6.1.1.1.1.80
TCDD and NBSA	83	Baseline spatial resolution	6.1.1.1.1.81
+	83	Baseline spatial resolution	6.1.1.1.1.82
TCDD and NBSA	84	Baseline spatial resolution	6.1.1.1.1.83
+	84	Baseline spatial resolution	6.1.1.1.1.84
TCDD and NBSA	85	Baseline spatial resolution	6.1.1.1.1.85
+	85	Baseline spatial resolution	6.1.1.1.1.86
TCDD and NBSA	86	Baseline spatial resolution	6.1.1.1.1.87
+	86	Baseline spatial resolution	6.1.1.1.1.88
TCDD and NBSA	87	Baseline spatial resolution	6.1.1.1.1.89
+	87	Baseline spatial resolution	6.1.1.1.1.90
TCDD and NBSA	88	Baseline spatial resolution	6.1.1.1.1.91
+	88	Baseline spatial resolution	6.1.1.1.1.92
TCDD and NBSA	89	Baseline spatial resolution	6.1.1.1.1.93
+	89	Baseline spatial resolution	6.1.1.1.1.94
TCDD and NBSA	90	Baseline spatial resolution	6.1.1.1.1.95
+	90	Baseline spatial resolution	6.1.1.1.1.96
TCDD and NBSA	91	Baseline spatial resolution	6.1.1.1.1.97
+	91	Baseline spatial resolution	6.1.1.1.1.98
TCDD and NBSA	92	Baseline spatial resolution	6.1.1.1.1.99
+	92	Baseline spatial resolution	6.1.1.1.1.100

21

## IPeM Criteria (example)

- Most results are subjective!

Reference paragraph	Physical parameter	Level of expertise	Frequency	Priority	Acceptable level	Suspension level
DOR01	Detector dose indicator monitoring	A	1-3 monthly	1	Baseline $\pm 20\%$	Baseline $\pm 50\%$ (a)
DOR02	Image uniformity	A	1-3 monthly	2	Lines or rectangles apparent	Lines non-uniformity
DOR03	Low contrast sensitivity	A	4-6 monthly	1	Baseline $\pm 20\%$	Baseline $\pm 50\%$ (a)
DOR04	Limiting spatial resolution	A	4-6 monthly	2	Baseline minus 25%	Baseline minus 25%
DOR05	Detector dose indicator repeatability	B	12 monthly	1	Baseline $\pm 20\%$	Baseline $\pm 50\%$ (a)
DOR06	Detector dose indicator reproducibility	B	12 monthly	1	Baseline $\pm 20\%$	Baseline $\pm 50\%$ (a)
DOR07	Motional blur/artifacts	B	12 monthly	1	Mean $\pm 5\%$	Mean $\pm 5\%$
DOR08	Threshold contrast (sharpness)	B	12 monthly	1	See Comments	See Comments
DOR09	Limiting spatial resolution	B	12 monthly	2	Baseline minus 25%	Baseline minus 25%
DOR10	Uniformity of exposure	B	12 monthly	2	Increase in blurring from baseline	Increase in blurring from baseline
DOR11	Scaling errors	B	12 monthly	2	$\pm 2\%$	$\pm 2\%$
DOR12	Dark noise	B	12 monthly	2	Baseline $\pm 50\%$	Baseline $\pm 50\%$

(a) These minimal and suspension levels are based on desirability, since the DOR is not linear with exposure.

22

## ACR Radiography/Fluoroscopy Accreditation Phantom

- Modules included
  - Chest
  - General
  - Fluoroscopy
- Phantom image
  - Radiography Chest/Abdomen
- ACR Discontinued Radiography/Fluoroscopy Accreditation Program in 2005




23

## Original Equipment Manufacturer (OEM) Products

- Automated Image Quality Control Tool
  - Reproducible quantitative results
  - May detect sub-visible changes in image quality performance to initiate timely preventive maintenance
  - Highly automated procedure
  - Most provide data reporting in spreadsheet format

24


## Test Phantom for Kodak (i.e. CareStream) DIRECTVIEW Total Quality Tool




\*Images provided by Eastman Kodak Company

## KODAK User Interface


- Uniformity
- Noise
- Spatial frequency response (MTF)
- Exposure linearity
- Pixel size accuracy and aspect ratio
- Phantom image artifacts
- Laser Beam Function
- Residual signal erase



## Kodak Test Result Details



## Kodak Temporal Test Results




## Test Limits

- Pre-set by OEM
- Basis for limit may not be justified in OEM literature
- If system fails a test, Service Engineer may not be educated how to correct problem

## DIN 6868-58 (2001) and 6868-13 (2002)

- Acceptance testing and constancy checks of projection radiography systems with digital image receptors
- German standard for testing of Storage Phosphor systems using a specially designed phantom to measure image quality parameters
- Can purchase a phantom that will meet the requirements of this standard from several vendors



### Anthropomorphic phantoms

- Shape "mimicking"




- Anatomically Accurate




31

### Shape "mimicking"




32

### Shape "mimicking"




33

### Anatomically Accurate



34

### Image Processing



35



### European Protocol for QC of ... mammography Addendum on Digital Mammography

- It is acknowledged that at present it is not possible to get unprocessed images from some systems.
- The image processing may introduce artifacts on phantom images and may be different from image processing for mammograms due to histogram or local texture based processing techniques.
- Therefore care needs to be taken in interpretation of these processed images.



36