

LINEAR ACCELERATOR QUALITY ASSURANCE TG-40 → TG-142

Eric E. Klein, Ph.D.
Chair of TG-142

Professor of Radiation Oncology
Washington University
St. Louis, MO

2011 AAPM Annual Meeting

AAPM Task Group 40 Report "Comprehensive QA for Radiation Oncology"

Med. Phys. 21(4) 1994

- Performance-based, comprehensive guidelines for preventing correctable systematic errors

- **Scope:**

- Guidelines for administrators
- Cobalt-60 Teletherapy Units
- Brachytherapy
- Conventional Simulators
- CT Scanners
- Measurement Equipment for Dosimetry
- Treatment Planning Computer Systems
- External Beam Treatment Planning Process
- External Beam QA for Individual Patients
- QA of Clinical Aspects

- QA of Medical Electron Accelerators → **Now TG-142**

Frequency	Parameter	Acceptance Criteria
Monthly	Dosimetry	
	x-ray output constancy ^a	2%
	Electron output constancy ^b	2%
	Backup monitor constancy	2%
	x-ray central axis dosimetry parameter (PDD, TAR) constancy	2%
	Electron central axis dosimetry parameter constancy (PDD)	2 mm @ therapeutic depth
	x-ray beam flatness constancy	2%
	Electron beam flatness constancy	3%
	x-ray and electron symmetry	3%
	Safety interlocks	Functional
	Emergency off switches	Functional
	Wedge, electron cone interlocks	Functional
	Mechanical Checks	
	Light/radiation field coincidence	2 mm or 1% on a side ^d
	Gantry/collimator angle indicators	1 deg
	Wedge position	2 mm (or 2% change in transmission factor)
	Tray position	2 mm
	Applimator position	2 mm
	Field size indicators	2 mm
	Cross-hair centering	2 mm diameter
Treatment couch position indicators	2 mm/1 deg	
Latching of wedges, blocking tray	Functional	
Jaw symmetry ^c	2 mm	
Field light intensity	Functional	

Monthly

TG-142: "QA of Medical Accelerators"

Med. Phys. 36(9) 2009

- Fills gap between TG-40 and TG-100
- Gives performance-based recommendations, but incorporates process-oriented concepts and advancements in linacs since 1994

- **Scope:** (replaces Table II of TG-40)

- Linac QA: acceptance testing, commissioning, CQI
- Ancillary treatment devices
 - Asymmetric jaws
 - Dynamic/virtual/universal wedge
 - MLC
 - TBI/TSET
 - Radiographic imaging
 - Respiratory gating

Task Group No. 100: Method for Evaluating QA Needs in Radiation Therapy

- Initially "Replacement for TG-40"
- Radical departure from previous AAPM recommendations and philosophy
- Based on "Failure Modes and Effects Analysis"
- Individual departments responsible for development of unique QA programs
- Based on procedures and resources performed at individual institutions

Task Group 100 Review e-mail

Dear QASC,

Cancel you vacations, TG 100 is here. All 188 pages, all 4700 lines, not to mention an alternate version chapter VII (itself only 114 pages). See Saifull's email below.

As you know, this much-awaited report has the potential to make a significant change to the way we do QA. We need to read this document carefully and consider how it can best be presented to allow implementation in the clinic. It is not just an educational document.

I would recommend you read through it the first time without trying to make any corrections or only ones you think are glaring. Get a feel for the whole thing. Then go back and re-read. I have read/skimmed most of it.

I would like us all to take about a month to read the document and be able to start to discuss. Let me know if you don't like that timeframe. After a month, I will set up a telecon for a general discussion. Don't worry about detailed line-by-line critique at this stage.

Art

Task Group 142: QA of Medical Accelerators

Members

- Chair: Eric E. Klein, Ph.D., Washington University
- Joseph Hanley, Ph.D., Hackensack Univ Medical Center
- John Bayouth, Ph.D., University of Iowa
- Fang-Fang Yin, Ph.D., Duke University
- William Simon, M.S., Sun Nuclear Corp.
- Sean Dresser, M.S., Northside Hospital
- Christopher Serago, Ph.D., Mayo Clinic, Jacksonville
- Francisco Aguirre, M.S., M.D. Anderson Cancer Center
- Lijun Ma, Ph.D., University of California, San Francisco
- Bijan Arjomandy, Ph.D., M.D. Anderson Cancer Center
- Chihray Liu, Ph.D., University of Florida
- Consultants: Carlos Sandin (Elekta), Todd Holmes (Varian Medical Systems)

Task Group 142: Philosophy

- The types of treatments delivered with the machine should also have a role in determining the QA program that is appropriate for that treatment machine.
- For example, machines that are used for SRS/SBRT treatments, TBI or IMRT require different tests and/or tolerances.


BACKGROUND

- Baseline dosimetric values entered into TPS to characterize and/or model the treatment machine directly affect calculated plans
- Values can deviate from their baseline as a result of;
 - Machine malfunction
 - Mechanical breakdown
 - Physical accidents
 - Component failure
 - Major component replacement
 - Gradual changes as a result of aging
- These patterns of failure must be considered when establishing a periodic QA program


TG-142 vs. TG-40

- TG-40 tests beam flatness/symmetry
 - A +/-3% drift in symmetry, while within TG-40 tolerance, means a 6% change in beam profile
 - New development: beams without flattening filters
- TG-142 recommends:
 - Beam profile measured with a QA device or portal imager
 - Several off-axis locations evaluated
 - Average of multiple points should be within tolerance values

Task Group 142: General



A Consistent beam profile is an important quantity for accurate and reproducible dose delivery in radiotherapy.



Copyright © 2010 Wolters Kluwer Health | Lippincott Williams & Wilkins


Task Group 142: General

Chosen O.A. points that fall within core of the field

$$\frac{1}{N} \cdot \sum_{L=1}^N \left| \frac{TP_L - BP_L}{BP_L} \right| \cdot 100\% \leq \text{Tolerance } \%$$

- where: TP_L and BP_L are off-axis ratios at Test and Baseline Points, respectively, at off axis Point L
- N is the number of off-axis points
- $TP_L = (MP_L / MP_C)$ where M represents the measured value, and C is the central axis measurement.
- Similarly, the baseline points are represented by $BP_L = (MBP_L / MBP_C)$

TG-142: Monthly



Light and Radiation Coincidence

Only needed if clinical setups using a light field is conducted

Copyright © 2010 Walters Kluwer Health | Lippincott Williams & Wilkins

TG-142 vs. TG-40


	Monthly	
	TG-40	TG-142 changes
Dosimetry		
x-ray central axis dosimetry parameter (PDD, TAR) constancy	2%	Removed
Electron central axis dosimetry parameter constancy (PDD)	2mm	2%/2mm
x-ray beam flatness constancy	2%	Replaced with 1% constancy of profile
Electron beam flatness constancy	3%	
x-ray and electron symmetry	3%	
Interlock Checks		
Emergency Off	Functional	Removed
Wedge, "cone"	Functional	
Mechanical		
Light/radiation field coincidence	2 mm or 1%/side	Only if clinical setups performed
Field size indicators	2mm	1mm/side
Cross-hair centering	2mm	1mm
Treatment couch position indicators	2 mm/1 deg	Tighter for SRS/SBRT

Annual

**If PDD₁₀₀ measured during TG51 calibration deviates >1%, discretion to measure more PDD points*

Procedure	Machine Type Tolerance		
	non-IMRT	IMRT	SRS/SBRT
Dosimetry			
X-ray flatness change from baseline		1%	
X-ray symmetry change from baseline		±1%	
Electron flatness change from baseline		1%	
Electron symmetry change from baseline		±1%	
SRS arc rotation mode (range: 0.5 to 10 MU/deg)	NA	NA	Monitor units set vs. delivered: 1.0 MU or 2% Gantry arc set vs. delivered: 1.0 deg or 2%
X-ray/electron output calibration (TG-51)	±1%(absolute)		
Spot check of field size dependent output factors for X-ray (2 or more FS)	2% for field size < 4x4 cm ² , 1% ≥4x4 cm ²		
Output factors for electron applicators (spot check of 1 applicator/energy)	±2% from baseline		
X-ray beam quality (PDD ₁₀ or TMR ₁₀ ²⁰)*	±1% from baseline		
Electron beam quality (R _p)	±1mm		

Annual: Mechanical & Radiation Isocenters




Copyright © 2010 Walters Kluwer Health | Lippincott Williams & Wilkins

Multileaf Collimation

Procedure	Tolerance
<i>Weekly (IMRT machines)</i>	
Qualitative test (i.e. matched segments, aka, "picket fence")	Visual inspection for discernable deviations such as an increase in interleaf transmission
<i>Monthly</i>	
Setting vs. radiation field for two patterns (non-IMRT)	2mm
Backup diaphragm settings (Elekta only)	2mm
Travel speed (IMRT)	Loss of leaf speed > 0.5 cm/sec
Leaf position accuracy (IMRT)	1mm for leaf positions of an IMRT field for 4 cardinal gantry angles. (Picket fence test may be used, test depends on clinical planning - segment size)
<i>Annually</i>	
MLC Transmission (Average of leaf and interleaf transmission), All Energies	±0.5% from baseline
Leaf position repeatability	±1.0 mm
MLC spoke shot	≤1.0 mm radius
Coincidence of Light Field and X-ray Field (All energies)	±2.0 mm
Segmental IMRT (Step and Shoot) Test	<0.35 cm Max Error RMS, 95% of error counts <0.35 cm
Moving window IMRT (4 cardinal gantry angles)	<0.35 cm Max Error RMS, 95% of error counts <0.35 cm

Multileaf Collimation: Weekly Qualitative *picket fence*



How to perform this test without a processor:
EPID or Radiochromic film


Imaging Tests: Daily

Or at a minimum when devices are to be used during treatment day

Procedure	Application Type Tolerance	
	non-SRS/SBRT	SRS/SBRT
Daily ¹¹		
kV and MV (EPID) imaging		
Collision interlocks	Functional	Functional
Positioning/repositioning	≤ 2 mm	≤ 1 mm
Imaging & Treatment coordinate coincidence (single gantry angle)	≤ 2 mm	≤ 1 mm
Cone-beam CT (kV & MV)		
Collision interlocks	Functional	Functional
Imaging & treatment coordinate coincidence	≤ 2 mm	≤ 1 mm
Positioning/repositioning	≤ 1 mm	≤ 1 mm


OBI QA

- Daily: Isocenter location and ability to shift accurately to a known location



OBI QA

Isocenter accuracy over gantry rotation



SUMMARY OF RECOMMENDATIONS/ IMPLEMENTATION SCHEME

- QA team led by the QMP supports all QA activities & policies and procedures.
- The 1st step is to establish institution-specific baseline and absolute reference values.
- Daily QA tasks may be carried out by a RTT using a cross-calibrated dosimetry system that is robust and easy-to-setup.
- There is overlap of tests for daily, monthly, and annual that can achieve independence with independent measurement devices.

SUMMARY OF RECOMMENDATIONS/ IMPLEMENTATION SCHEME

- End-to-end system checks ensure fidelity of overall system.
- During the annual QA, absolute outputs should be calibrated as per TG51 and all secondary QA dosimeters cross-checked.
- Upon completion of an annual QA report be generated, signed and reviewed by the QMP and filed for future machine maintenance and inspection needs.