

## Erratum: “Report of AAPM TG 135: Quality assurance for robotic radiosurgery”

Sonja Dieterich

*Stanford University Cancer Center, Stanford, California 94305*

Carlo Cavedon

*Azienda Ospedaliera Universitaria Integrata di Verona, U.O. di Fisica Sanitaria, Verona 37126, Italy*

Cynthia F. Chuang

*Department of Radiation Oncology, University of California San Francisco, San Francisco, California 94143-0226*

Alan B. Cohen

*Accuray Inc., Sunnyvale, California 94089*

Jeffrey A. Garrett

*Mississippi Baptist Medical Center, Jackson, Mississippi 39202*

Charles L. Lee

*CK Solutions, Inc., Edmond, Oklahoma 73034*

Jessica R. Lowenstein

*UT MD Anderson Cancer Center, Houston, Texas 77030*

Maximian F. d'Souza

*St Anthony Hospital, Oklahoma City, Oklahoma 73101*

David D. Taylor, Jr.

*US Radiosurgery, Nashville, Tennessee 80304*

Xiaodong Wu

*Department of Radiation Oncology, University of Miami, Miami, Florida 33101*

Cheng Yu

*USC Keck School of Medicine, Los Angeles, California 90033*

(Received 26 July 2011; accepted for publication 27 July 2011; published 30 August 2011)

[DOI: [10.1118/1.3626480](https://doi.org/10.1118/1.3626480)]

We would like to correct a sentence on p. 2932, Sec. III E. General Patient Safety, *Treatment Procedure Monitoring*. The incorrect sentence reads: “We recommend that in addition to the treating therapist, a second medical professional (therapist, physicist, or physician) should be in the immediate vicinity at all times to assist when necessary.

The corrected section follows below. We regret the error and hope the revised text will clarify the task groups recommendations in regards to the role of QMPs in treatment supervision.

Glossary:

Personal supervision—Physicist is present at the treatment console.

Direct supervision—Physicist is in the department, available for immediate response.

Corrected text:

All treatments must occur under direct supervision of a QMP. In addition, a QMP must provide personal supervision at the first treatment, and as needed for subsequent treatments. The personal supervision should include participation in a time-out checklist, assessment of patient immobilization, assessment of adequate imaging parameters, tracking accuracy and correct Synchrony respiratory modeling (if applicable), consultation on excessive or unusual patient shift requirements during treatment not clearly caused by patient motion on the treatment couch, as well as other patient- or plan-specific needs.